


## SABRE Medical Oxygen Regulators

Sabre Medical Oxygen Regulators must be serviced **every five years**. Details for the next date of service can be found on the waist label. (see the diagram below ). Cylinder Test Period and Regulator Service must be kept in line.

### Example



---

**PLEASE CHECK THE SERVICE DUE DATE  
ON YOUR REGULATOR**